One sample preparation for all mycotoxins

The only B1 product for Baby Food Water or Methanol Extraction kits for Deoxynivalenol (DON)

Very Fast Protocol (5 mins)

No Purification Columns needed

Ochratoxin Fumonisin Zearalenone Deoxynivalenol

Aflatoxin B1 Aflatoxin Total T-2 / HT-2

MYCOTOXIN DETERMINATION IN GRAINS, NUTS, CEREALS & ANIMAL FEED

Exports to 40 countries globally
ELISA kits for all regulated mycotoxins
Very low LOQ
12 months self-life
All reagents in ready-to-use format

FAPAS FOOD CHEMISTRY PROFICIENCY TESTS, 2016-2019

Multi-mycotoxins in Maize (Corn)	December 2016 Z-score	September 2017 Z-score	March 2018 Z-score	March 2019 Z-score	April-May 2019 Z-score	Average Z-score
Aflatoxin B1 (AFB1)	0.5	0.4	-0.5	-0.3	0.2	0.06
Deoxynivalenol	0	-0.1	0.5	-0.6	-0.6	-0.16
Zearalenone (ZON)	0.1	0.1	-0.1	-0.2	0.0	-0.02
Ochratoxin A (OTA)	0.4	0.6	0	0.5	0.1	0.32

BIBLIOGRAPHY

Chatzoglou Ch, Natsaridis N, Dimopoulos M and Athana-

siou SD, An innovate rapid two-incubation immunoassay to optimize the recovery levels of ochratoxin A in overparticular matrices by dilution normalization. IAFP's European Symposium on Food Safety. 11-13 May 2016, Athens, Greece. (Poster presentation)

Chatzoglou Ch, Natsaridis N, Manouras A, Malissiova E and Ntantasios A, Optimization of total aflatoxin recovery levels in overparticular matrice s using a twenty-minute ELISA by dilution normalization and two incubation steps.

ELISA by dilution normalization and two incubation steps. IAFP's European Symposium on Food Safety. 11-13 May 2016, Athens, Greece. (Poster presentation)

Ntantasios A, Papageorgiou G, Chatzoglou Ch and Athanasiou SD, Optimization of recovery levels of total aflatoxin in overparticular matrices in immunoassays by dilution normalization. 38th Mycotoxin Workshop. 2-4 May, 2016, Berlin, Germany. (Poster presentation)

PROGNOSIS BIOTECH S.A.

FARSALON 153 | 41335 LARISSA-GREECE e: info@prognosis-biotech.com web: www.prognosis-biotech.com

